
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 005292
13:44 28/03/2023
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:03
PRESSURIZING 3:04
STERILIZING 3:59
136.0 C 220kPa 3:59
136.3 C 223kPa 4:29
136.4 C 225kPa 4:59
136.2 C 221kPa 5:29
136.4 C 223kPa 5:59
136.4 C 225kPa 6:29
136.3 C 221kPa 6:59
136.4 C 223kPa 7:29
136.4 C 223kPa 7:29
VENTING 7:29
AIR DRYING 8:23
CYCLE COMPLETE 10:51

Digital Signature #
C778A8FC514A313F
