

-----  
Midmark M11 - v1.0.5  
Total Cycles: 4406

Name: RITTER2  
Sterilizer ID: V2313228  
03/06/2025 12:16 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 96.6 F 0.0  
2:00 97.0 F 0.5  
4:00 104.2 F 1.8  
6:00 129.5 F 4.3  
8:00 161.0 F 7.3  
10:00 184.3 F 10.6  
12:00 210.1 F 7.3  
14:00 210.1 F 8.0  
16:00 223.9 F 12.1  
18:00 228.4 F 8.0  
20:00 238.8 F 13.0  
22:00 238.7 F 11.0  
24:00 249.3 F 16.6  
26:00 260.2 F 22.5  
28:00 269.8 F 28.9

STERILIZING  
mm:ss Degrees PSI  
0:00 271.8 F 30.7  
1:00 271.8 F 30.6  
2:00 271.8 F 30.6  
3:00 272.1 F 30.6  
4:00 272.4 F 30.8

Min 271.5 F 30.5  
Max 272.7 F 31.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 205.5 F 0.0  
5:00 201.9 F 0.0  
10:00 207.8 F 0.0  
15:00 202.8 F 0.0  
20:00 200.8 F 0.0  
25:00 202.7 F 0.0  
30:00 203.5 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 29:38  
STERILIZING: 4:00  
VENTING: 3:51  
DRYING: 30:00  
TOTAL CYCLE: 01:08:22  
CYCLE COMPLETE

03/06/2025 01:24 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_