
Midmark M11 - v1.0.5
Total Cycles: 2323

Name: RITTER2
Sterilizer ID: V2313228
05/02/2023 08:59 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 136.9 F 0.0
2:00 137.0 F 1.5
4:00 147.3 F 4.2
6:00 172.2 F 8.2
8:00 213.3 F 13.1
10:00 215.2 F 7.6
12:00 228.9 F 13.0
14:00 231.0 F 9.6
16:00 236.5 F 9.4
18:00 244.7 F 13.9
20:00 256.5 F 20.3
22:00 267.6 F 27.2

STERILIZING
mm:ss Degrees PSI
0:00 272.2 F 30.0
1:00 272.1 F 30.5
2:00 272.7 F 30.5
3:00 271.7 F 29.9
4:00 272.5 F 30.8

Min 271.7 F 29.9
Max 272.9 F 30.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.4 F 0.0
5:00 198.5 F 0.0
10:00 208.5 F 0.0
15:00 211.3 F 0.0
20:00 222.7 F 0.0
25:00 215.9 F 0.0
30:00 222.0 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 23:16
STERILIZING: 4:00
VENTING: 3:16
DRYING: 30:00
TOTAL CYCLE: 01:01:32
CYCLE COMPLETE

05/02/2023 09:59 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____