

-----  
Herzig Eye Institute Ottawa Inc  
phone/email  
150 Bloor St. West, Suite 210  
Toronto ON M5S2X9  
North York, ON M5S2X9  
Canada

6472616554  
david@herzig-eye.com  
-----

STATIM 2000 S203R751  
SN 500919I00001  
Unit # : 000  
WATER QUALITY  
0.2uS / 0.1ppm  
CYCLE NUMBER 007490  
11:33 04/06/2024  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:24  
PRESSURIZING 3:25  
STERILIZING 4:29  
136.4 C 223kPa 4:29  
136.4 C 223kPa 4:59  
136.4 C 223kPa 5:29  
136.4 C 223kPa 5:59  
136.2 C 223kPa 6:29  
136.0 C 224kPa 6:59  
136.3 C 223kPa 7:29  
136.4 C 223kPa 7:59  
136.4 C 223kPa 7:59  
VENTING 7:59  
AIR DRYING 8:52  
CYCLE COMPLETE 10:36

-----  
Digital Signature #  
FBEDA534C53F0366  
-----