
Midmark M11 - v1.0.0
Total Cycles: 4923

Name: RITTER 1
Sterilizer ID: V2178198
09/19/2024 02:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 103.9 F 0.0
2:00 111.5 F 0.0
4:00 137.4 F 0.0
6:00 165.1 F 0.3
8:00 188.0 F 1.6
10:00 203.9 F 4.9
12:00 215.1 F 8.4
14:00 226.3 F 12.4
16:00 228.9 F 7.4
18:00 240.2 F 12.1
20:00 237.2 F 9.5
22:00 239.2 F 9.6
24:00 247.7 F 14.0
26:00 259.0 F 20.1
28:00 269.2 F 26.6

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.8
2:00 272.5 F 28.8
3:00 271.9 F 28.3
4:00 271.5 F 28.2

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 188.6 F 0.0
5:00 203.4 F 0.0
10:00 207.2 F 0.0
15:00 193.4 F 0.0
20:00 218.3 F 0.0
25:00 223.2 F 0.0
30:00 225.9 F 0.0
Drying Complete

FILLING: 0:51
HEATING: 29:14
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:49
CYCLE COMPLETE

09/19/2024 03:54 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____