
Midmark M11 - v1.0.5
Total Cycles: 1955

Name: RITTER2
Sterilizer ID: V2313228
01/23/2023 09:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	76.3 F	0.0
2:00	82.4 F	0.0
4:00	124.4 F	0.0
6:00	109.1 F	0.0
8:00	141.0 F	0.2
10:00	169.3 F	0.5
12:00	192.9 F	0.9
14:00	206.7 F	1.6
16:00	216.3 F	2.4
18:00	221.5 F	3.2
20:00	224.6 F	4.2
22:00	229.6 F	6.0
24:00	232.7 F	7.2
26:00	235.6 F	8.4
28:00	243.0 F	11.8
30:00	235.5 F	8.5
32:00	245.6 F	13.1
34:00	240.8 F	10.8
36:00	236.6 F	8.5
38:00	246.7 F	13.7

C533: HEATUP MODE
STEAM TEMP LOW
CHAMBER TEMP 259F
CHAMBER PRESS 20.6PSI
** ITEMS NOT STERILE **

01/23/2023 10:10 AM