
Midmark M11 - v1.0.0
Total Cycles: 3605

Name: RITTER 1
Sterilizer ID: V2178198
08/02/2023 08:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 71.1 F 0.0
2:00 73.0 F 0.0
4:00 79.0 F 0.0
6:00 107.3 F 0.0
8:00 141.7 F 0.2
10:00 177.2 F 0.9
12:00 192.2 F 4.2
14:00 204.2 F 7.3
16:00 215.6 F 10.5
18:00 226.6 F 9.1
20:00 231.6 F 8.3
22:00 239.9 F 12.6
24:00 237.7 F 9.4
26:00 239.9 F 9.7
28:00 247.0 F 13.4
30:00 258.5 F 19.6
32:00 268.7 F 26.0

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.5
2:00 271.7 F 28.3
3:00 272.5 F 28.4
4:00 271.6 F 28.2

Min 271.4 F 27.8
Max 272.7 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 197.7 F 0.0
5:00 208.2 F 0.0
10:00 201.1 F 0.0
15:00 230.3 F 0.0
20:00 231.4 F 0.0
25:00 231.6 F 0.0
30:00 227.0 F 0.0

Drying Complete

FILLING: 0:47
HEATING: 33:13
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:11:16
CYCLE COMPLETE

08/02/2023 09:41 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____