
Midmark M11 - v1.0.5
Total Cycles: 1145

Name: RITTER2
Sterilizer ID: V2313228
05/31/2022 11:08 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.2 F 0.0
2:00 194.6 F 4.0
4:00 203.3 F 9.7
6:00 217.8 F 6.2
8:00 232.0 F 10.8
10:00 233.5 F 8.4
12:00 241.0 F 10.6
14:00 247.4 F 14.5
16:00 259.8 F 21.4
18:00 270.9 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 29.1
1:00 272.7 F 29.7
2:00 271.6 F 29.0
3:00 272.9 F 29.8
4:00 272.0 F 29.1

Min 271.6 F 28.9
Max 273.0 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 203.2 F 0.0
5:00 203.1 F 0.0
10:00 215.5 F 0.0
15:00 214.7 F 0.0
20:00 218.8 F 0.0
25:00 229.8 F 0.0
30:00 206.4 F 0.0
Drying Complete

FILLING: 2:17
HEATING: 18:45
STERILIZING: 4:00
VENTING: 3:08
DRYING: 30:00
TOTAL CYCLE: 00:58:11
CYCLE COMPLETE

05/31/2022 12:10 PM
ACCEPTED: X
REJECTED:
Operator: U02