

-----  
Midmark M11 - v1.0.0  
Total Cycles: 973

Name: RITTER 1  
Sterilizer ID: V2178198  
07/07/2021 08:47 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 121.0 F 0.0  
2:00 128.5 F 0.0  
4:00 128.8 F 0.5  
6:00 166.1 F 2.7  
8:00 188.5 F 5.9  
10:00 207.9 F 9.7  
12:00 225.4 F 10.5  
14:00 227.0 F 7.2  
16:00 237.9 F 11.8  
18:00 235.2 F 8.6  
20:00 242.4 F 11.3  
22:00 244.6 F 12.5  
24:00 256.5 F 18.7  
26:00 267.2 F 25.1

STERILIZING  
mm:ss Degrees PSI  
0:00 271.1 F 28.2  
1:00 272.7 F 28.7  
2:00 271.2 F 28.3  
3:00 272.6 F 28.7  
4:00 271.1 F 28.2

Min 271.1 F 28.1  
Max 272.7 F 28.9

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.2 F 0.0  
5:00 232.8 F 0.0  
10:00 224.9 F 0.0  
15:00 225.1 F 0.0  
20:00 230.1 F 0.0  
25:00 225.4 F 0.0  
30:00 220.0 F 0.0

Drying Complete

FILLING: 0:50  
HEATING: 27:33  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:05:40  
CYCLE COMPLETE

07/07/2021 09:59 AM  
ACCEPTED: X  
REJECTED:  
Operator: U02