
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 007893
10:19 26/09/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:56
PRESSURIZING 2:02
STERILIZING 3:06
136.4 C 222kPa 3:06
135.9 C 221kPa 3:36
136.4 C 225kPa 4:06
136.4 C 225kPa 4:36
136.4 C 222kPa 5:06
136.4 C 223kPa 5:36
136.4 C 225kPa 6:06
136.4 C 225kPa 6:36
136.4 C 225kPa 6:36
VENTING 6:36
AIR DRYING 7:28
CYCLE COMPLETE 15:11

Digital Signature #
03BA86AFC2E880A8
