
Midmark M11 - v1.0.0
Total Cycles: 3900

Name: RITTER 1
Sterilizer ID: V2178198
10/13/2023 10:31 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 199.4 F 0.0
2:00 191.6 F 3.5
4:00 201.8 F 8.2
6:00 219.6 F 10.3
8:00 229.7 F 8.7
10:00 241.6 F 11.3
12:00 241.8 F 11.8
14:00 240.2 F 10.4
16:00 253.9 F 17.3
18:00 265.5 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.8 F 28.3
1:00 272.3 F 29.1
2:00 271.9 F 28.4
3:00 271.9 F 28.9
4:00 272.3 F 28.7

Min 271.3 F 28.1
Max 272.5 F 29.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.6 F 0.0
5:00 224.4 F 0.0
10:00 233.1 F 0.0
15:00 226.9 F 0.0
20:00 226.7 F 0.0
25:00 228.2 F 0.0
30:00 223.5 F 0.0

Drying Complete

FILLING: 2:05
HEATING: 19:36
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 00:59:06
CYCLE COMPLETE

10/13/2023 11:30 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____