

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3623

Name: RITTER 1  
Sterilizer ID: V2178198  
08/09/2023 08:40 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

mm:ss	Degrees	PSI
0:00	69.8 F	0.0
2:00	70.2 F	0.0
4:00	76.8 F	0.0
6:00	105.5 F	0.0
8:00	134.1 F	0.2
10:00	163.3 F	1.8
12:00	186.8 F	4.7
14:00	204.0 F	8.1
16:00	210.2 F	11.7
18:00	217.9 F	4.8
20:00	225.5 F	9.0
22:00	235.1 F	13.0
24:00	234.4 F	9.3
26:00	239.7 F	10.3
28:00	244.2 F	12.5
30:00	255.2 F	18.2
32:00	265.1 F	24.1
34:00	271.2 F	28.3

STERILIZING

mm:ss	Degrees	PSI
0:00	271.2 F	28.4
1:00	272.5 F	29.2
2:00	272.1 F	28.6
3:00	271.3 F	28.4
4:00	272.6 F	29.2

Min 271.2 F 28.2  
Max 272.6 F 29.2

VENTING CHAMBER

DRYING

Drying Start

mm:ss	Degrees	PSI
0:00	192.1 F	0.0
5:00	202.9 F	0.0
10:00	238.8 F	0.0
15:00	229.3 F	0.0
20:00	219.1 F	0.0
25:00	232.1 F	0.0
30:00	232.5 F	0.0

Drying Complete

FILLING: 0:47  
HEATING: 34:04  
STERILIZING: 4:00  
VENTING: 3:16  
DRYING: 30:00  
TOTAL CYCLE: 01:12:07  
CYCLE COMPLETE

08/09/2023 09:52 AM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_