
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001182
14:10 21/07/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:47
PRESSURIZING 2:56
STERILIZING 4:09
136.5 C 224kPa 4:09
136.1 C 223kPa 4:39
136.4 C 225kPa 5:09
136.4 C 225kPa 5:39
136.4 C 223kPa 6:09
136.2 C 224kPa 6:39
136.1 C 222kPa 7:09
136.3 C 224kPa 7:39
136.3 C 224kPa 7:39
VENTING 7:39
AIR DRYING 8:31
CYCLE COMPLETE 21:34

Digital Signature #
48A0FBE43BF194E0
