
Midmark M11 - v1.0.0
Total Cycles: 3826

Name: RITTER 1
Sterilizer ID: V2178198
09/27/2023 09:49 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 176.6 F 0.0
2:00 168.8 F 0.0
4:00 171.5 F 1.0
6:00 192.6 F 3.9
8:00 208.1 F 8.2
10:00 222.2 F 12.8
12:00 228.6 F 8.1
14:00 239.1 F 13.3
16:00 241.7 F 11.0
18:00 238.3 F 9.3
20:00 252.1 F 15.9
22:00 263.2 F 22.3
24:00 272.4 F 28.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.2
1:00 272.2 F 28.5
2:00 272.2 F 28.8
3:00 272.0 F 28.3
4:00 272.4 F 28.9

Min 271.4 F 28.1
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.9 F 0.0
5:00 232.1 F 0.0
10:00 203.2 F 0.0
15:00 204.1 F 0.0
20:00 194.1 F 0.0
25:00 203.2 F 0.0
30:00 240.2 F 0.0
Drying Complete

FILLING: 0:49
HEATING: 24:57
STERILIZING: 4:00
VENTING: 3:32
DRYING: 30:00
TOTAL CYCLE: 01:03:18
CYCLE COMPLETE

09/27/2023 10:52 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____