
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000543
11:43 09/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:29
PRESSURIZING 2:34
STERILIZING 3:41
136.3 C 220kPa 3:41
136.4 C 233kPa 4:11
136.3 C 233kPa 4:41
136.2 C 233kPa 5:11
136.4 C 233kPa 5:41
136.4 C 233kPa 6:11
136.2 C 231kPa 6:41
136.2 C 231kPa 7:11
136.2 C 231kPa 7:11
VENTING 7:11
AIR DRYING 8:05
CYCLE COMPLETE 23:05

Digital Signature #
FE87466605B2DA09
