
Midmark M11 - v1.0.0
Total Cycles: 1404

Name: RITTER 1
Sterilizer ID: V2178198
11/30/2021 07:54 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 72.8 F 0.0
2:00 89.1 F 0.0
4:00 142.6 F 0.0
6:00 107.3 F 0.1
8:00 157.5 F 0.1
10:00 176.1 F 0.3
12:00 204.5 F 1.8
14:00 217.1 F 4.9
16:00 224.9 F 7.3
18:00 233.6 F 10.4
20:00 238.2 F 9.9
22:00 237.9 F 10.0
24:00 238.8 F 9.1
26:00 245.6 F 13.0
28:00 242.8 F 11.5
30:00 255.1 F 17.7
32:00 265.9 F 24.2

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.7 F 28.6
2:00 271.6 F 28.0
3:00 272.7 F 28.7
4:00 271.5 F 27.9

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 211.7 F 0.0
5:00 234.8 F 0.0
10:00 195.0 F 0.0
15:00 185.4 F 0.0
20:00 170.5 F 0.0
25:00 168.8 F 0.0
30:00 171.3 F 0.0

Drying Complete

FILLING: 1:01
HEATING: 33:45
STERILIZING: 4:00
VENTING: 3:05
DRYING: 30:00
TOTAL CYCLE: 01:11:50
CYCLE COMPLETE

11/30/2021 09:13 AM
ACCEPTED: X
REJECTED:
Operator: U02