
Brampton, ON
Canada

STATIM 5000 S501R713
SN 511019D00004
Unit # : 000
WATER QUALITY
1.0uS / 0.6ppm
CYCLE NUMBER 001483
9:39 10/07/2021
UNWRAPPED
134 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 4:34
PRESSURIZING 6:20
STERILIZING 7:54
136.0 C 220kPa 7:54
136.4 C 222kPa 8:24
135.8 C 217kPa 8:54
136.3 C 221kPa 9:24
136.4 C 221kPa 9:54
136.3 C 222kPa 10:24
136.3 C 221kPa 10:54
136.4 C 223kPa 11:24
VENTING 11:24
AIR DRYING 12:28
CYCLE COMPLETE 21:00

Digital Signature #
704987FB98A373F6
