
Midmark M11 - v1.0.0
Total Cycles: 5199

Name: RITTER 1
Sterilizer ID: V2178198
12/17/2024 03:09 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 107.2 F 0.0
2:00 112.1 F 0.0
4:00 137.8 F 0.0
6:00 178.9 F 0.1
8:00 197.8 F 0.2
10:00 215.4 F 1.8
12:00 227.0 F 5.4
14:00 233.5 F 8.9
16:00 243.8 F 13.0
18:00 239.2 F 10.1
20:00 238.3 F 9.1
22:00 245.9 F 12.4
24:00 245.7 F 12.8
26:00 257.7 F 19.2
28:00 268.1 F 25.7

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.0
1:00 272.4 F 28.6
2:00 272.7 F 28.7
3:00 272.2 F 28.3
4:00 271.7 F 27.9

Min 271.5 F 27.8
Max 272.8 F 28.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 181.6 F 0.0
5:00 244.3 F 0.0
10:00 198.4 F 0.0
15:00 194.4 F 0.0
20:00 180.3 F 0.0
25:00 178.6 F 0.0
30:00 176.7 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 29:28
STERILIZING: 4:00
VENTING: 3:41
DRYING: 30:00
TOTAL CYCLE: 01:07:54
CYCLE COMPLETE

12/17/2024 04:16 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____