
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.1uS / 0.0ppm
CYCLE NUMBER 004404
9:07 21/09/2022

UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:43
PRESSURIZING 1:47
STERILIZING 2:51
136.4 C 222kPa 2:51
136.4 C 223kPa 3:21
136.2 C 221kPa 3:51
136.4 C 223kPa 4:21
136.4 C 222kPa 4:51
136.1 C 221kPa 5:21
136.4 C 224kPa 5:51
136.2 C 221kPa 6:21
136.2 C 221kPa 6:21
VENTING 6:21
AIR DRYING 7:17
CYCLE COMPLETE 8:20

Digital Signature #
F451C7D2FFAE5861
