
Midmark M11 - v1.0.5
Total Cycles: 1653

Name: RITTER2
Sterilizer ID: V2313228
10/25/2022 11:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 153.4 F 0.0
2:00 156.4 F 1.7
4:00 166.4 F 5.1
6:00 194.6 F 10.3
8:00 224.8 F 5.1
10:00 229.7 F 10.5
12:00 234.5 F 7.3
14:00 241.3 F 12.8
16:00 240.6 F 11.3
18:00 253.3 F 17.8
20:00 264.7 F 24.6

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 29.2
1:00 271.4 F 29.6
2:00 272.5 F 29.8
3:00 271.5 F 29.3
4:00 271.1 F 29.3

Min 271.0 F 29.1
Max 272.7 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 209.4 F 0.0
5:00 194.4 F 0.0
10:00 220.9 F 0.0
15:00 231.3 F 0.0
20:00 227.4 F 0.0
25:00 224.3 F 0.0
30:00 225.7 F 0.0
Drying Complete

FILLING: 1:43
HEATING: 21:47
STERILIZING: 4:00
VENTING: 3:10
DRYING: 30:00
TOTAL CYCLE: 01:00:41
CYCLE COMPLETE

10/25/2022 12:52 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____