
Midmark M11 - v1.0.5
Total Cycles: 1630

Name: RITTER2
Sterilizer ID: V2313228
10/20/2022 10:28 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 187.7 F 0.0
2:00 176.9 F 2.8
4:00 186.5 F 8.2
6:00 214.7 F 8.7
8:00 230.9 F 9.0
10:00 236.7 F 10.1
12:00 243.0 F 12.8
14:00 242.5 F 12.0
16:00 256.0 F 19.1
18:00 267.6 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.0
1:00 272.4 F 29.7
2:00 271.2 F 29.1
3:00 272.4 F 29.5
4:00 271.2 F 29.1

Min 271.2 F 28.9
Max 272.6 F 29.8

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 191.8 F 0.0
5:00 233.2 F 0.0
10:00 230.9 F 0.0
15:00 221.4 F 0.0
20:00 232.1 F 0.0
25:00 227.2 F 0.0
30:00 191.0 F 0.0
Drying Complete

FILLING: 2:20
HEATING: 19:09
STERILIZING: 4:00
VENTING: 3:15
DRYING: 30:00
TOTAL CYCLE: 00:58:44
CYCLE COMPLETE

10/20/2022 11:26 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____