
Midmark M11 - v1.0.0
Total Cycles: 1318

Name: RITTER 1
Sterilizer ID: V2178198
10/29/2021 06:52 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 73.1 F 0.0
2:00 77.9 F 0.0
4:00 91.1 F 0.0
6:00 119.2 F 0.0
8:00 151.9 F 0.5
10:00 179.5 F 2.9
12:00 194.8 F 5.9
14:00 208.4 F 9.2
16:00 220.3 F 12.8
18:00 225.3 F 6.5
20:00 236.6 F 10.9
22:00 233.3 F 7.4
24:00 245.6 F 13.5
26:00 243.8 F 12.0
28:00 256.2 F 18.5
30:00 267.2 F 25.3

STERILIZING
mm:ss Degrees PSI
0:00 272.0 F 28.2
1:00 272.7 F 29.0
2:00 271.6 F 28.2
3:00 272.5 F 28.7
4:00 272.0 F 28.6

Min 271.5 F 28.1
Max 272.7 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 204.5 F 0.0
5:00 213.1 F 0.0
10:00 208.3 F 0.0
15:00 207.9 F 0.0
20:00 201.9 F 0.0
25:00 196.7 F 0.0
30:00 192.7 F 0.0
Drying Complete

FILLING: 0:56
HEATING: 31:18
STERILIZING: 4:00
VENTING: 3:06
DRYING: 30:00
TOTAL CYCLE: 01:09:20
CYCLE COMPLETE

10/29/2021 08:01 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____