
Midmark M11 - v1.0.0
Total Cycles: 4099

Name: RITTER 1
Sterilizer ID: V2178198
12/08/2023 02:40 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 104.5 F 0.0
2:00 115.2 F 1.0
4:00 126.6 F 3.3
6:00 157.8 F 6.3
8:00 180.6 F 9.7
10:00 195.2 F 13.5
12:00 208.3 F 5.5
14:00 220.1 F 9.5
16:00 235.4 F 10.5
18:00 235.4 F 9.8
20:00 237.7 F 9.3
22:00 245.4 F 13.1
24:00 256.5 F 19.0
26:00 266.9 F 25.4

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.6
1:00 271.7 F 29.0
2:00 272.6 F 29.1
3:00 271.4 F 28.5
4:00 271.5 F 28.9

Min 271.1 F 28.4
Max 272.7 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 189.0 F 0.0
5:00 216.2 F 0.0
10:00 219.1 F 0.0
15:00 219.9 F 0.0
20:00 218.1 F 0.0
25:00 219.0 F 0.0
30:00 217.5 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 27:29
STERILIZING: 4:00
VENTING: 3:40
DRYING: 30:00
TOTAL CYCLE: 01:06:08
CYCLE COMPLETE

12/08/2023 03:46 PM
ACCEPTED:
REJECTED:

Reason: _____
Operator: _____