

-----  
Midmark M11 - v1.0.0  
Total Cycles: 3545

Name: RITTER 1  
Sterilizer ID: V2178198  
07/18/2023 01:31 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 208.5 F 0.0  
2:00 206.9 F 3.8  
4:00 218.1 F 9.1  
6:00 224.6 F 7.6  
8:00 234.8 F 10.3  
10:00 236.9 F 7.8  
12:00 242.3 F 11.4  
14:00 246.2 F 13.4  
16:00 258.9 F 20.2  
18:00 270.3 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.6 F 28.3  
1:00 272.5 F 28.9  
2:00 271.6 F 28.4  
3:00 272.6 F 29.0  
4:00 271.8 F 28.3  
  
Min 271.5 F 28.2  
Max 272.7 F 29.2

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.6 F 0.0  
5:00 226.4 F 0.0  
10:00 231.7 F 0.0  
15:00 229.1 F 0.0  
20:00 227.6 F 0.0  
25:00 228.9 F 0.0  
30:00 230.2 F 0.0  
Drying Complete

FILLING: 2:13  
HEATING: 18:50  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:58:17  
CYCLE COMPLETE

07/18/2023 02:29 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_