
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.3uS / 0.1ppm
CYCLE NUMBER 000553
13:50 23/07/2021
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:30
PRESSURIZING 2:38
STERILIZING 3:51
136.0 C 228kPa 3:51
135.9 C 229kPa 4:21
136.1 C 229kPa 4:51
136.4 C 230kPa 5:21
136.3 C 232kPa 5:51
136.3 C 232kPa 6:21
136.4 C 232kPa 6:51
136.4 C 233kPa 7:21
136.4 C 233kPa 7:21
VENTING 7:21
AIR DRYING 8:15
CYCLE COMPLETE 23:15

Digital Signature #
81B14433B0CE1817
