
Midmark M11 - v1.0.0
Total Cycles: 5421

Name: RITTER 1
Sterilizer ID: V2178198
03/19/2025 08:19 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 68.7 F 0.0
2:00 72.0 F 0.2
4:00 84.0 F 1.0
6:00 123.6 F 3.4
8:00 147.3 F 6.0
10:00 171.4 F 9.1
12:00 198.6 F 12.6
14:00 203.6 F 3.8
16:00 215.1 F 7.2
18:00 225.8 F 10.8
20:00 227.5 F 6.9
22:00 237.7 F 10.6
24:00 234.4 F 7.7
26:00 247.2 F 14.1
28:00 258.8 F 20.3
30:00 269.1 F 26.9

STERILIZING
mm:ss Degrees PSI
0:00 271.6 F 28.5
1:00 272.2 F 28.6
2:00 272.6 F 28.9
3:00 272.8 F 29.3
4:00 272.0 F 28.8

Min 271.6 F 28.4
Max 272.8 F 29.3

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 186.9 F 0.0
5:00 201.5 F 0.0
10:00 189.6 F 0.0
15:00 192.2 F 0.0
20:00 179.4 F 0.0
25:00 195.3 F 0.0
30:00 200.2 F 0.0
Drying Complete

FILLING: 0:44
HEATING: 31:01
STERILIZING: 4:00
VENTING: 3:25
DRYING: 30:00
TOTAL CYCLE: 01:09:11
CYCLE COMPLETE

03/19/2025 09:28 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____