
Herzig Eye Institute Ottawa Inc
phone/email
150 Bloor St. West, Suite 210
Toronto ON M5S2X9
North York, ON M5S2X9
Canada

6472616554
david@herzig-eye.com

STATIM 2000 S203R751
SN 500919I00001
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 008106
15:19 27/11/2024
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 2:02
PRESSURIZING 2:06
STERILIZING 3:09
136.4 C 222kPa 3:09
136.3 C 222kPa 3:39
136.1 C 221kPa 4:09
136.4 C 224kPa 4:39
136.4 C 224kPa 5:09
136.4 C 223kPa 5:39
136.3 C 220kPa 6:09
136.1 C 219kPa 6:39
136.1 C 219kPa 6:39
VENTING 6:39
AIR DRYING 7:29
CYCLE COMPLETE 11:35

Digital Signature #
7028972065CC50B3
