

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 001375  
10:04 02/12/2022  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:55  
PRESSURIZING 3:58  
STERILIZING 5:22  
136.0 C 222kPa 5:22  
136.0 C 221kPa 5:52  
136.3 C 223kPa 6:22  
136.4 C 224kPa 6:52  
136.3 C 224kPa 7:22  
136.2 C 224kPa 7:52  
136.3 C 222kPa 8:22  
136.4 C 224kPa 8:52  
136.4 C 224kPa 8:52  
VENTING 8:52  
AIR DRYING 9:50  
CYCLE COMPLETE 24:50  
-----

Digital Signature #  
C6BF8633089123D7  
-----