
Midmark M11 - v1.0.0
Total Cycles: 3209

Name: RITTER 1
Sterilizer ID: V2178198
05/10/2023 09:04 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 194.4 F 0.0
2:00 184.0 F 0.0
4:00 183.6 F 3.6
6:00 217.1 F 8.3
8:00 221.8 F 13.6
10:00 230.7 F 8.4
12:00 242.4 F 13.4
14:00 239.7 F 10.9
16:00 237.3 F 9.1
18:00 251.0 F 15.7
20:00 262.5 F 22.3
22:00 272.6 F 28.7

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 28.2
1:00 272.4 F 28.6
2:00 271.7 F 28.7
3:00 272.0 F 28.4
4:00 271.8 F 28.8

Min 271.2 F 28.1
Max 272.6 F 29.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 210.5 F 0.0
5:00 236.3 F 0.0
10:00 227.0 F 0.0
15:00 227.8 F 0.0
20:00 213.3 F 0.0
25:00 211.7 F 0.0
30:00 209.0 F 0.0
Drying Complete

FILLING: 1:24
HEATING: 22:18
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 01:00:60
CYCLE COMPLETE

05/10/2023 10:04 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____