
Midmark M11 - v1.0.5
Total Cycles: 1004

Name: RITTER2
Sterilizer ID: V2313228
04/19/2022 10:30 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 213.6 F 0.0
2:00 203.7 F 3.1
4:00 209.2 F 9.2
6:00 225.5 F 8.2
8:00 236.2 F 11.9
10:00 237.2 F 10.3
12:00 237.4 F 9.5
14:00 251.5 F 16.7
16:00 263.7 F 23.9
18:00 271.5 F 29.2

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.2
1:00 272.4 F 29.5
2:00 272.0 F 29.8
3:00 272.0 F 29.3
4:00 272.1 F 30.0

Min 271.4 F 29.1
Max 272.6 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 213.4 F 0.0
5:00 230.6 F 0.0
10:00 229.6 F 0.0
15:00 228.3 F 0.0
20:00 228.2 F 0.0
25:00 229.8 F 0.0
30:00 227.6 F 0.0

Drying Complete

FILLING: 2:38
HEATING: 18:00
STERILIZING: 4:00
VENTING: 3:14
DRYING: 30:00
TOTAL CYCLE: 00:57:52
CYCLE COMPLETE

04/19/2022 11:27 AM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____