
Midmark M11 - v1.0.5
Total Cycles: 2480

Name: RITTER2
Sterilizer ID: V2313228
06/07/2023 12:27 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 198.8 F 0.0
2:00 195.5 F 4.0
4:00 199.5 F 9.2
6:00 219.9 F 7.4
8:00 231.2 F 11.0
10:00 233.5 F 8.9
12:00 237.6 F 9.7
14:00 248.8 F 15.2
16:00 261.6 F 22.5
18:00 272.8 F 29.8

STERILIZING
mm:ss Degrees PSI
0:00 271.4 F 29.2
1:00 272.2 F 29.6
2:00 271.8 F 29.7
3:00 272.1 F 29.5
4:00 271.7 F 29.7

Min 271.3 F 29.1
Max 272.7 F 30.0

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.8 F 0.0
5:00 221.6 F 0.0
10:00 236.3 F 0.0
15:00 229.2 F 0.0
20:00 227.3 F 0.0
25:00 226.3 F 0.0
30:00 228.3 F 0.0
Drying Complete

FILLING: 1:58
HEATING: 18:18
STERILIZING: 4:00
VENTING: 3:17
DRYING: 30:00
TOTAL CYCLE: 00:57:34
CYCLE COMPLETE

06/07/2023 01:24 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____