
Midmark M11 - v1.0.5
Total Cycles: 3156

Name: RITTER2
Sterilizer ID: V2313228
11/22/2023 04:47 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 131.4 F 0.0
2:00 134.3 F 1.5
4:00 155.6 F 4.8
6:00 182.6 F 9.3
8:00 205.7 F 7.0
10:00 221.1 F 8.6
12:00 231.5 F 10.8
14:00 238.5 F 11.7
16:00 238.4 F 10.5
18:00 252.6 F 17.5
20:00 264.9 F 24.8

STERILIZING
mm:ss Degrees PSI
0:00 271.9 F 29.3
1:00 272.5 F 30.1
2:00 271.7 F 29.2
3:00 272.4 F 30.1
4:00 272.0 F 29.4

Min 271.2 F 29.1
Max 272.8 F 30.1

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 193.5 F 0.0
5:00 213.2 F 0.0
10:00 208.0 F 0.0
15:00 202.9 F 0.0
20:00 201.5 F 0.0
25:00 200.3 F 0.0
30:00 200.3 F 0.0
Drying Complete

FILLING: 0:59
HEATING: 21:38
STERILIZING: 4:00
VENTING: 3:19
DRYING: 30:00
TOTAL CYCLE: 00:59:56
CYCLE COMPLETE

11/22/2023 05:46 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____