

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
1.0uS / 0.6ppm  
CYCLE NUMBER 000308  
9:55 20/11/2020  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:25  
PRESSURIZING 2:33  
STERILIZING 4:01  
136.5 C 232kPa 4:01  
136.2 C 232kPa 4:31  
136.3 C 232kPa 5:01  
136.2 C 232kPa 5:31  
136.3 C 233kPa 6:01  
136.3 C 232kPa 6:31  
136.4 C 233kPa 7:01  
136.4 C 233kPa 7:31  
136.4 C 233kPa 7:31  
VENTING 7:31  
AIR DRYING 8:25  
CYCLE COMPLETE 11:00  
-----

Digital Signature #  
626E1C18B8A278C5  
-----