
Midmark M11 - v1.0.0
Total Cycles: 3967

Name: RITTER 1
Sterilizer ID: V2178198
11/01/2023 02:13 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 99.1 F 0.0
2:00 100.7 F 0.9
4:00 120.1 F 3.0
6:00 149.9 F 5.9
8:00 176.8 F 9.3
10:00 212.5 F 13.5
12:00 216.5 F 4.9
14:00 226.4 F 8.4
16:00 236.0 F 12.3
18:00 233.0 F 8.0
20:00 243.9 F 12.8
22:00 241.2 F 10.7
24:00 252.4 F 16.3
26:00 262.6 F 22.3
28:00 272.5 F 28.8

STERILIZING
mm:ss Degrees PSI
0:00 271.3 F 28.2
1:00 272.4 F 28.5
2:00 271.8 F 28.7
3:00 272.0 F 28.4
4:00 272.0 F 28.8

Min 271.2 F 28.1
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 199.8 F 0.0
5:00 238.2 F 0.0
10:00 222.8 F 0.0
15:00 218.3 F 0.0
20:00 193.9 F 0.0
25:00 218.8 F 0.0
30:00 237.4 F 0.0
Drying Complete

FILLING: 1:01
HEATING: 28:24
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:07:10
CYCLE COMPLETE

11/01/2023 03:20 PM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____