
Midmark M11 - v1.0.5
Total Cycles: 702

Name: RITTER2
Sterilizer ID: V2313228
01/17/2022 01:50 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 184.5 F 0.0
2:00 179.8 F 3.6
4:00 188.9 F 8.0
6:00 211.4 F 13.3
8:00 225.8 F 9.2
10:00 234.9 F 9.8
12:00 241.8 F 13.2
14:00 242.8 F 12.6
16:00 255.9 F 19.5
18:00 267.4 F 26.7

STERILIZING
mm:ss Degrees PSI
0:00 271.0 F 29.5
1:00 271.4 F 29.6
2:00 271.9 F 30.0
3:00 271.2 F 29.7
4:00 271.5 F 29.7

Min 271.0 F 29.4
Max 272.1 F 30.2

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 192.2 F 0.0
5:00 233.4 F 0.0
10:00 215.5 F 0.0
15:00 202.9 F 0.0
20:00 199.1 F 0.0
25:00 198.1 F 0.0
30:00 197.3 F 0.0
Drying Complete

FILLING: 1:55
HEATING: 19:18
STERILIZING: 4:00
VENTING: 3:09
DRYING: 30:00
TOTAL CYCLE: 00:58:22
CYCLE COMPLETE

01/17/2022 02:48 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____