
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.4uS / 0.2ppm
CYCLE NUMBER 001033
14:02 08/04/2022
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 3:09
PRESSURIZING 3:14
STERILIZING 4:25
136.4 C 224kPa 4:25
136.4 C 225kPa 4:55
136.2 C 224kPa 5:25
136.2 C 223kPa 5:55
136.3 C 225kPa 6:25
136.4 C 224kPa 6:55
136.3 C 224kPa 7:25
136.3 C 224kPa 7:55
136.3 C 224kPa 7:55
VENTING 7:55
AIR DRYING 8:49
CYCLE COMPLETE 23:49

Digital Signature #
C5E4F72364D300D7
