
Midmark M11 - v1.0.0
Total Cycles: 1445

Name: RITTER 1
Sterilizer ID: V2178198
12/09/2021 03:54 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 197.5 F 0.0
2:00 194.4 F 3.8
4:00 204.5 F 8.7
6:00 222.4 F 8.5
8:00 233.5 F 9.4
10:00 236.7 F 9.2
12:00 245.2 F 13.0
14:00 243.8 F 11.9
16:00 257.5 F 19.0
18:00 269.2 F 26.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.0
1:00 272.7 F 28.6
2:00 271.5 F 27.9
3:00 272.7 F 28.8
4:00 271.7 F 28.0

Min 271.4 F 27.8
Max 272.8 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 194.6 F 0.0
5:00 214.8 F 0.0
10:00 219.4 F 0.0
15:00 211.8 F 0.0
20:00 178.0 F 0.0
Dry Cycle Aborted

FILLING: 2:50
HEATING: 19:02
STERILIZING: 4:00
VENTING: 3:04
DRYING: 21:49
TOTAL CYCLE: 00:50:46

12/09/2021 04:44 PM
ACCEPTED: X
REJECTED:
Operator: U02