

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.3uS / 0.1ppm  
CYCLE NUMBER 000343  
12:37 15/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 3:01  
PRESSURIZING 3:06  
STERILIZING 4:17  
136.5 C 232kPa 4:17  
136.3 C 233kPa 4:47  
136.4 C 233kPa 5:17  
136.4 C 233kPa 5:47  
136.3 C 232kPa 6:17  
136.2 C 232kPa 6:47  
136.4 C 233kPa 7:17  
136.2 C 230kPa 7:47  
136.2 C 230kPa 7:47  
VENTING 7:47  
AIR DRYING 8:39  
CYCLE COMPLETE 23:39  
-----

Digital Signature #  
525D72F721BA3AA4  
-----