
Midmark M11 - v1.0.0
Total Cycles: 3380

Name: RITTER 1
Sterilizer ID: V2178198
06/13/2023 10:40 AM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 191.4 F 0.0
2:00 188.8 F 1.5
4:00 192.2 F 6.1
6:00 212.6 F 11.0
8:00 221.9 F 6.4
10:00 235.2 F 11.8
12:00 234.6 F 9.1
14:00 238.6 F 9.5
16:00 247.2 F 14.1
18:00 259.5 F 20.9
20:00 270.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 28.6
1:00 272.6 F 29.2
2:00 271.5 F 28.5
3:00 272.6 F 29.3
4:00 271.9 F 28.7

Min 271.5 F 28.5
Max 272.6 F 29.4

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 212.8 F 0.0
5:00 219.4 F 0.0
10:00 232.2 F 0.0
15:00 229.9 F 0.0
20:00 225.7 F 0.0
25:00 225.7 F 0.0
30:00 223.3 F 0.0
Drying Complete

FILLING: 1:12
HEATING: 20:48
STERILIZING: 4:00
VENTING: 3:18
DRYING: 30:00
TOTAL CYCLE: 00:59:18
CYCLE COMPLETE

06/13/2023 11:39 AM
ACCEPTED:
REJECTED:
Reason: _____
Operator: _____