

-----  
Midmark M11 - v1.0.0  
Total Cycles: 5370

Name: RITTER 1  
Sterilizer ID: V2178198  
02/26/2025 02:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 216.4 F 0.0  
2:00 210.0 F 4.4  
4:00 224.3 F 9.8  
6:00 228.7 F 7.0  
8:00 235.0 F 11.3  
10:00 236.2 F 9.0  
12:00 240.0 F 9.7  
14:00 248.1 F 14.4  
16:00 260.0 F 20.8  
18:00 269.9 F 27.5

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.7 F 29.6  
2:00 272.5 F 29.0  
3:00 271.5 F 28.7  
4:00 272.4 F 29.6  
  
Min 271.5 F 28.6  
Max 272.9 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 214.3 F 0.0  
5:00 217.6 F 0.0  
10:00 231.7 F 0.0  
15:00 230.9 F 0.0  
20:00 230.1 F 0.0  
25:00 228.1 F 0.0  
30:00 228.1 F 0.0  
Drying Complete

FILLING: 3:18  
HEATING: 19:04  
STERILIZING: 4:00  
VENTING: 3:53  
DRYING: 30:00  
TOTAL CYCLE: 01:00:14  
CYCLE COMPLETE

02/26/2025 03:49 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_