
North York, ON
Canada

STATIM 2000 S203R750
SN 500919I00007
Unit # : 000
WATER QUALITY
0.2uS / 0.1ppm
CYCLE NUMBER 002431
11:56 17/01/2025
UNWRAPPED
135 C for 3.5 min.
CYCLE START 0:00
CONDITIONING 1:55
PRESSURIZING 2:03
STERILIZING 3:05
136.4 C 223kPa 3:05
135.8 C 221kPa 3:35
136.4 C 223kPa 4:05
136.4 C 223kPa 4:35
136.2 C 224kPa 5:05
136.4 C 226kPa 5:35
136.2 C 223kPa 6:05
136.4 C 225kPa 6:35
136.4 C 225kPa 6:35
VENTING 6:35
AIR DRYING 7:28
CYCLE COMPLETE 37:28

Digital Signature #
16C9D58EDC34EC49
