

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4868

Name: RITTER 1  
Sterilizer ID: V2178198  
09/05/2024 08:41 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 69.3 F 0.0  
2:00 71.1 F 0.0  
4:00 84.9 F 0.0  
6:00 118.9 F 0.2  
8:00 152.5 F 0.6  
10:00 179.1 F 0.8  
12:00 198.5 F 3.1  
14:00 211.7 F 6.2  
16:00 220.4 F 9.8  
18:00 231.3 F 11.3  
20:00 232.8 F 8.6  
22:00 243.9 F 13.3  
24:00 240.9 F 10.8  
26:00 236.6 F 8.6  
28:00 250.8 F 15.2  
30:00 261.6 F 21.3  
32:00 271.4 F 27.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 27.8  
1:00 272.6 F 28.4  
2:00 271.5 F 27.9  
3:00 272.5 F 28.4  
4:00 271.4 F 27.7  
  
Min 271.4 F 27.6  
Max 272.7 F 28.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 178.2 F 0.0  
5:00 193.8 F 0.0  
10:00 188.5 F 0.0  
15:00 193.4 F 0.0  
20:00 228.1 F 0.0  
25:00 231.7 F 0.0  
30:00 229.5 F 0.0  
  
Drying Complete

FILLING: 0:41  
HEATING: 32:39  
STERILIZING: 4:00  
VENTING: 3:20  
DRYING: 30:00  
TOTAL CYCLE: 01:10:40  
CYCLE COMPLETE

09/05/2024 09:51 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_