

-----  
Midmark M11 - v1.0.5  
Total Cycles: 748

Name: RITTER2  
Sterilizer ID: V2313228  
02/01/2022 04:49 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 170.3 F 0.0  
2:00 176.3 F 3.6  
4:00 201.3 F 9.0  
6:00 217.9 F 7.9  
8:00 231.7 F 10.0  
10:00 234.9 F 8.5  
12:00 239.5 F 10.1  
14:00 249.9 F 15.4  
16:00 262.6 F 22.7  
18:00 272.2 F 28.8

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.8  
1:00 272.5 F 29.2  
2:00 271.7 F 29.0  
3:00 272.4 F 29.2  
4:00 271.5 F 28.7  
  
Min 271.5 F 28.6  
Max 272.7 F 29.6

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 181.5 F 0.0  
5:00 231.5 F 0.0  
10:00 212.9 F 0.0  
15:00 210.7 F 0.0  
20:00 212.6 F 0.0  
25:00 215.5 F 0.0  
30:00 219.3 F 0.0  
Drying Complete

FILLING: 2:08  
HEATING: 18:09  
STERILIZING: 4:00  
VENTING: 3:07  
DRYING: 30:00  
TOTAL CYCLE: 00:57:24  
CYCLE COMPLETE

02/01/2022 05:46 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_