

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.1uS / 0.0ppm  
CYCLE NUMBER 002585  
10:00 21/03/2025  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 1:51  
PRESSURIZING 1:57  
STERILIZING 3:08  
136.4 C 224kPa 3:08  
136.3 C 225kPa 3:38  
136.2 C 221kPa 4:08  
136.3 C 224kPa 4:38  
136.4 C 226kPa 5:08  
136.4 C 225kPa 5:38  
136.4 C 226kPa 6:08  
136.3 C 226kPa 6:38  
136.3 C 226kPa 6:38  
VENTING 6:38  
AIR DRYING 7:29  
CYCLE COMPLETE 37:29

-----  
Digital Signature #  
BD5A7911FB5D9930  
-----