
Midmark M11 - v1.0.0
Total Cycles: 5148

Name: RITTER 1
Sterilizer ID: V2178198
12/02/2024 03:41 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 87.3 F 0.0
2:00 87.8 F 0.0
4:00 109.3 F 0.0
6:00 136.9 F 0.1
8:00 154.5 F 0.3
10:00 179.6 F 0.6
12:00 197.4 F 1.3
14:00 210.3 F 4.6
16:00 219.7 F 7.9
18:00 230.3 F 11.5
20:00 229.0 F 6.8
22:00 241.6 F 12.0
24:00 237.8 F 9.4
26:00 240.0 F 9.8
28:00 246.8 F 13.4
30:00 257.2 F 18.9
32:00 266.2 F 24.5
34:00 271.6 F 28.0

STERILIZING
mm:ss Degrees PSI
0:00 271.7 F 28.1
1:00 271.9 F 28.4
2:00 272.5 F 28.9
3:00 272.6 F 28.8
4:00 272.3 F 28.5

Min 271.5 F 28.0
Max 272.7 F 28.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 182.8 F 0.0
5:00 217.0 F 0.0
10:00 227.9 F 0.0
15:00 242.9 F 0.0
20:00 234.6 F 0.0
25:00 229.4 F 0.0
30:00 224.0 F 0.0
Drying Complete

FILLING: 0:45
HEATING: 34:03
STERILIZING: 4:00
VENTING: 3:45
DRYING: 30:00
TOTAL CYCLE: 01:12:33
CYCLE COMPLETE

12/02/2024 04:53 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____