

-----  
Midmark M11 - v1.0.0  
Total Cycles: 2580

Name: RITTER 1  
Sterilizer ID: V2178198  
11/17/2022 05:06 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 203.6 F 0.0  
2:00 201.0 F 4.7  
4:00 203.9 F 9.9  
6:00 218.5 F 5.8  
8:00 233.5 F 10.9  
10:00 234.2 F 8.2  
12:00 240.9 F 10.4  
14:00 247.7 F 14.1  
16:00 260.7 F 21.3  
18:00 272.5 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.4 F 28.3  
1:00 272.3 F 28.8  
2:00 271.5 F 28.4  
3:00 272.4 F 28.9  
4:00 271.5 F 28.2  
  
Min 271.4 F 28.2  
Max 272.5 F 29.1

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 213.7 F 0.0  
5:00 236.4 F 0.0  
10:00 230.8 F 0.0  
15:00 226.9 F 0.0  
20:00 227.0 F 0.0  
25:00 228.1 F 0.0  
30:00 228.0 F 0.0  
Drying Complete

FILLING: 1:51  
HEATING: 18:25  
STERILIZING: 4:00  
VENTING: 3:14  
DRYING: 30:00  
TOTAL CYCLE: 00:57:31  
CYCLE COMPLETE

11/17/2022 06:03 PM

ACCEPTED:

REJECTED:

Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_