
Midmark M11 - v1.0.5
Total Cycles: 1339

Name: RITTER2
Sterilizer ID: V2313228
07/25/2022 01:23 PM

BEGIN POUCHES CYCLE
Temp: 270 Degrees F
Time: 4 Minutes
Dry: 30 Minutes

FILLING CHAMBER

HEATING
mm:ss Degrees PSI
0:00 196.1 F 0.0
2:00 195.4 F 1.8
4:00 202.0 F 7.5
6:00 220.2 F 13.2
8:00 231.4 F 10.2
10:00 233.2 F 8.3
12:00 239.3 F 10.4
14:00 248.1 F 14.9
16:00 260.7 F 21.9
18:00 271.9 F 29.4

STERILIZING
mm:ss Degrees PSI
0:00 271.5 F 29.1
1:00 272.3 F 29.6
2:00 271.5 F 29.2
3:00 272.4 F 29.7
4:00 271.7 F 29.0

Min 271.4 F 28.9
Max 272.6 F 29.9

VENTING CHAMBER

DRYING
Drying Start
mm:ss Degrees PSI
0:00 206.5 F 0.0
5:00 220.3 F 0.0
10:00 236.3 F 0.0
15:00 229.8 F 0.0
20:00 225.6 F 0.0
25:00 223.5 F 0.0
30:00 221.9 F 0.0
Drying Complete

FILLING: 1:52
HEATING: 18:33
STERILIZING: 4:00
VENTING: 3:11
DRYING: 30:00
TOTAL CYCLE: 00:57:36
CYCLE COMPLETE

07/25/2022 02:20 PM

ACCEPTED:

REJECTED:

Reason: _____
Operator: _____