

-----  
Midmark M11 - v1.0.5  
Total Cycles: 1361

Name: RITTER2  
Sterilizer ID: V2313228  
08/03/2022 08:56 AM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 150.9 F 0.0  
2:00 152.1 F 0.0  
4:00 169.3 F 0.2  
6:00 182.7 F 2.1  
8:00 202.8 F 6.4  
10:00 220.3 F 11.0  
12:00 226.2 F 6.7  
14:00 237.1 F 12.2  
16:00 237.2 F 10.1  
18:00 236.8 F 8.7  
20:00 249.8 F 15.4  
22:00 261.2 F 21.8  
24:00 271.9 F 29.0

STERILIZING  
mm:ss Degrees PSI  
0:00 271.2 F 28.6  
1:00 272.4 F 29.5  
2:00 272.2 F 29.1  
3:00 271.4 F 28.5  
4:00 271.6 F 29.1

Min 271.2 F 28.5  
Max 272.5 F 29.5

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 203.7 F 0.0  
5:00 223.7 F 0.0  
10:00 230.1 F 0.0  
15:00 226.7 F 0.0  
20:00 224.5 F 0.0  
25:00 235.7 F 0.0  
30:00 227.6 F 0.0  
Drying Complete

FILLING: 0:53  
HEATING: 24:37  
STERILIZING: 4:00  
VENTING: 3:08  
DRYING: 30:00  
TOTAL CYCLE: 01:02:39  
CYCLE COMPLETE

08/03/2022 09:58 AM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_