

-----  
Midmark M11 - v1.0.0  
Total Cycles: 4778

Name: RITTER 1  
Sterilizer ID: V2178198  
07/31/2024 01:09 PM

BEGIN POUCHES CYCLE  
Temp: 270 Degrees F  
Time: 4 Minutes  
Dry: 30 Minutes

FILLING CHAMBER

HEATING  
mm:ss Degrees PSI  
0:00 97.7 F 0.0  
2:00 98.3 F 0.9  
4:00 117.7 F 3.2  
6:00 152.2 F 6.2  
8:00 179.3 F 9.7  
10:00 203.9 F 6.7  
12:00 209.8 F 5.5  
14:00 222.0 F 9.5  
16:00 234.2 F 11.0  
18:00 234.9 F 10.0  
20:00 235.8 F 8.6  
22:00 246.0 F 13.5  
24:00 257.4 F 19.6  
26:00 268.1 F 26.2

STERILIZING  
mm:ss Degrees PSI  
0:00 271.5 F 28.7  
1:00 272.6 F 29.2  
2:00 271.4 F 28.6  
3:00 272.6 F 29.3  
4:00 271.8 F 28.7

Min 271.4 F 28.5  
Max 272.8 F 29.4

VENTING CHAMBER

DRYING  
Drying Start  
mm:ss Degrees PSI  
0:00 188.0 F 0.0  
5:00 202.9 F 0.0  
10:00 198.6 F 0.0  
15:00 208.8 F 0.0  
20:00 234.7 F 0.0  
25:00 229.4 F 0.0  
30:00 225.4 F 0.0  
Drying Complete

FILLING: 0:46  
HEATING: 27:21  
STERILIZING: 4:00  
VENTING: 3:32  
DRYING: 30:00  
TOTAL CYCLE: 01:05:39  
CYCLE COMPLETE

07/31/2024 02:14 PM  
ACCEPTED:  
REJECTED:  
Reason: \_\_\_\_\_  
Operator: \_\_\_\_\_