

-----  
North York, ON  
Canada  
-----

STATIM 2000 S203R750  
SN 500919I00007  
Unit # : 000  
WATER QUALITY  
0.4uS / 0.2ppm  
CYCLE NUMBER 000364  
15:43 20/01/2021  
UNWRAPPED  
135 C for 3.5 min.  
CYCLE START 0:00  
CONDITIONING 2:29  
PRESSURIZING 2:37  
STERILIZING 3:35  
136.4 C 232kPa 3:35  
135.9 C 230kPa 4:05  
136.2 C 231kPa 4:35  
136.1 C 230kPa 5:05  
136.1 C 230kPa 5:35  
136.3 C 233kPa 6:05  
136.4 C 231kPa 6:35  
136.2 C 231kPa 7:05  
136.2 C 231kPa 7:05  
VENTING 7:05  
AIR DRYING 7:58  
CYCLE COMPLETE 8:16  
-----

Digital Signature #  
44F22099A04523BF  
-----